

DONGSHENG BIOTECH CO., LTD.

Room 305, Building A, No. 179, Guangpu East Road, Huangpu Distric,
Guangzhou, Guangdong, China

CERTIFICATE OF ANALYSIS

Product Name:	SYBR Green qPCR Mix	According to:	Enterprise Standard
Batch No.:	104	Mfg. Date:	2025/4/9
Exp. Date:	2027/4/8	Report Date:	2025/4/9
Storage:	-20°C		

Specifications	Results
At least 3 of 4 NTC (No Template Control) reactions must yield Ct values ≥ 35	Pass
Average Ct for 10 copies of human genomic DNA must be ≤ 35	Pass
The correlation coefficient for the standard curve must be ≥ 0.99	Pass
The slope of the standard curve must be between -3.2 and - 3.8	Pass

Conclusion: Qualified

QC Supervisor: Sen Lu

Checker: Louis Wang

Analyst: May Chan



DONGSHENG BIOTECH CO., LTD.

Room 305, Building A, No. 179, Guangpu East Road, Huangpu Distric,
Guangzhou, Guangdong, China

CERTIFICATE OF ANALYSIS

Product Name:	Nuclease-free water	According to:	Enterprise Standard
Batch No.:	058	Mfg. Date:	2025/9/16
Exp. Date:	2027/9/15	Report Date:	2025/9/16
Storage:	-20°C		

Testing Items	Specifications	Results
Appearance	Clear colorless solution	Clear colorless solution
Stability	The Water, nuclease-free, is deionized and 0.22 µm membrane-filtered. It is suitable for use in all molecular biology applications	Pass

Conclusion: Qualified

QC Supervisor: Sen Lu

Checker: Louis Wang

Analyst: May Chan

