

# DONGSHENG BIOTECH CO., LTD.

Room 305, Building A, No. 179, Guangpu East Road, Huangpu District,  
Guangzhou, Guangdong, China

## CERTIFICATE OF ANALYSIS

Product Name:	ARMS PCR Mix	According to:	Enterprise Standard
Cat. No.:	P4011	Mfg. Date:	2024/10/26
Batch No.:	1005	Exp. Date:	2026/10/25
Storage:	-20°C	Report Date:	2024/10/26

Testing Items	Specifications	Results
Appearance	Clear colorless solution	Clear colorless solution
Endonuclease residue of the stock solution of enzymes	Not detected in 2000 U	Not detected
Exonuclease residue of the stock solution of enzymes	Not detected in 2000 U	Not detected
RNase residue of the stock solution of enzymes	Not detected in 2000 U	Not detected
Functional testing	Consistent amplification was observed in real-time PCR using $\lambda$ DNA as a template. No specific amplification was observed without template.	Pass

**Conclusion: Qualified**

QC Supervisor: WZ Mai

Checker: TL Nie

Analyst: QQ Chen

